

InertSearch™ for LC

Inertsil® Applications

Diphenhydramine in Control Serum

Data No.LA157-0000



Column : Bioptic AV-2
150 × 4.6mm I.D.

Solutes : 1. Diphenhydramine

Column Cat. No. : 5020-02512

Eluent : 50mM KH₂PO₄(pH2.3)/CH₃CN=95/5

Flow rate : 1.0 mL/min.

Col. Temp. : 30°C

Detection : UV 210 nm

Injection Vol. :

Sample : Diphenhydramine