



# iReceptal

## Surgical Suction Management

- Enhanced infection control
- Heightened clinical efficacy
- Optimised cost management



# iReceptal

## Surgical Suction Management

### Waste fluid collection

- One-way filter (4 ports) provides multiple suction ports
- 20L waste fluid capacity is optimal for all surgical procedures
- Specimen trap, a single use accessory, separates tissues from the waste fluid, preventing any potential back flow

### Automatic coupling, waste discharge and self cleaning

- The drainage and cleaning cycle activates automatically when the Rover is docked with the docking station
- Docking station releases the Rover automatically after the cleaning cycle


### Suction Rover mobility

- Unique design of the Rover allows for ease of transport and location transfer
- Completely closed system

## iReceptal compared to traditional treatments

Product			
Automatic discharge	No	No	Yes
Cleaning	Manual cleaning	-	Automatic cleaning
Turnover time	20-30 minutes	20 minutes	5-6 minutes
Volume	2-3 litres	1-2 litres	20litres

# Innovative Features



- Completely closed system for easy use.
- Collects and disposes of large volumes of surgical waste fluid (20L).
- Helps optimise O.R. turnaround time, scheduling, and volume.
- Includes integrated tissue trap for specimen collection.
- Provides economic advantages alongside clinical benefits.
- Sustainable solution: reduces plastic and clinical bag waste in every fluid generating department.

## Product specifications:

Model Specifications	iRR100-01 iReceptal Rover 1.0	iRR150-01 iReceptal Rover 1.5	iRD102-01 iReceptal Docking Station
Height	105 cm (41 inches)	105 cm (41 inches)	55 cm (22 inches)
Depth	51 cm (20 inches)	51 cm (20 inches)	56 cm (22 inches)
Width	105 cm (41 inches)	105 cm (41 inches)	55 cm (22 inches)
Volume	48 cm (19 inches)	48 cm (19 inches)	63 cm (25 inches)
Power supply	Battery 4.5V, AAx3	48 cm (19 inches)	220-230V, 50/60Hz, 350W
Weight	32 kg (71 lbs.)	34 kg (75 lbs.)	45 kg (99 lbs.)
Plumbing requirements	Docking Station should be plumbed to a water connection with a minimum of 45 PSI.		

## Product order information:

Reference	Description
iRR100-01	iReceptal Rover 1.0 (International)
iRR150-01	iReceptal Rover 1.5 (International)
iRD102-01	iReceptal Docking Station (220V)
iRD101-01	iReceptal Docking Station (110V)
iRM400-01	iReceptal 4-Port Manifold
IRC001-01	iReceptal Docking Station Detergent
IRC002-01	iReceptal Docking Station Detergent Concentrate
VGK001	VAC=GARD Kit (Vacuum Guard Kit)

Visit our website:

[www.fannin.eu](http://www.fannin.eu)



Fannin (UK) Ltd  
Unit 3 Fern House  
Daniels Industrial  
Estate  
104 Bath Road  
Stroud GL5 3TJ