

A Non-hazardous Technique to Determine Penicillin G and Penicillin G Procaine in Beef


Data No. LL003-0000

First published in *LCGC Europe*, 22(7), 354-356, (2009)

Provided by Naoto Furusawa, an associate professor with the Graduate School of Human Life Science, Osaka City University, Osaka 558-8585, Japan.

Author name : Naoto Furusawa
Data source : *LCGC Europe*, 22(7), 354-356, (2009)
Year : 2009

Figure 1: Chromatograms obtained from the HPLC system. PDA detector set at (a) 205 nm or (b) 290 nm. Upper profile: a beef (cattle muscle) sample spiked with the target compounds (0.2 µg/g). Lower profile: (a) a blank beef sample. Peaks: PG (Retention time) = 2.4 min; 2 = PGp (17.3 min). An arrow in (a) lower profile indicates the peak of PG.



Conditions

Column : Inertsil WP 300 C4 (5 µm, 100 x 4.6 mm I.D.)
Column Cat. No. : 5020-05894
Eluent : A) 0.04 mol/L phosphoric acid (pH 7.0)
 : B) Ethanol
 : A/B = 8/2, v/v
Flow Rate : 1.0 mL/min
Col. Temp. : 40 °C
Detection : PDA 190-350 nm (detected at 205 and 290 nm)
Injection Vol. : 20 µL
Sample : beef
Analyte : 1. Penicillin G (PG)
 : 2. PG procaine (PGp)
URL : <http://chromatographyonline.findanalytichem.com/lcgc/article/articleDetail.jsp?id=610621&pageID=1&sk=&date=> (last access date: Jan. 6, 2010)