


## Analysis of Fluocinolone acetonide(column select)

Data No. LA171-0000



$R_s=2.5$

Column :Inertsil ODS-3V  
5 $\mu$ m  
250  $\times$  4.6mmI.D.

Solutes : 1. iso-Propyl paraben(0.05mg/mL)  
2. n-Propyl paraben(0.05mg/mL)

Column Cat. No. :

Eluent :CH<sub>3</sub>CN/H<sub>2</sub>O=30/70

Flow rate :2.0 mL/min.

Col. Temp. :40°C

Detection :UV 254 nm

Injection Vol. :20 $\mu$ L

Sample : 1. iso-Propyl paraben(0.05mg/mL)  
2. n-Propyl paraben(0.05mg/mL)