


## Analysis of Fluocinolone acetonide

Data No. LA182-0000



Column : Inertsil SIL-100A  
5 $\mu$ m  
150  $\times$  4.6mm I.D.

Solutes : 1. Triamcinolone acetonide(1.5mg/mL)  
2. Fluocinolone acetonide(1.5mg/mL)

Column Cat. No. :

Eluent : CHCl<sub>3</sub>(aq.saturated)/CH<sub>3</sub>OH/CH<sub>3</sub>COOH  
=1000/15/10

Flow rate : 1.0 mL/min.

Col. Temp. : 30°C

Detection : UV 254 nm

Injection Vol. : 20 $\mu$ L

Sample : 1. Triamcinolone acetonide(1.5mg/mL)  
2. Fluocinolone acetonide(1.5mg/mL)