


InertSearch™ for LC

Inertsil™ Applications

Analysis of Glutathione

Data No. LB577-0806



Conditions

System : GL7700 HPLC system
Column : InertSustain PFP HP (3 μ m, 150 x 2.1 mm I.D.)
Column Cat. No. : 5020-87921
Eluent : 0.1% HCOOH in H₂O
Flow rate : 0.2 mL/min
Col. Temp. : 40 °C
Detection : UV 230 nm (UV7750 with semi-micro flow cell)
Injection Vol. : 1 μ L
Sample : Standard

Analyte:

1. GSH (Reduced form of Glutathione)
500 mg/L
2. GSSG (Oxidized form of Glutathione)
500 mg/L