


Analysis of Glycyrrhiza Standard (Under the Condition of Japanese Pharmacopoeia)

Data No. LA550-0698



System	: GL-7400 system	Solutes	: 1. Glycyrrhizic acid dipotassium salt (0.27 mg/mL)
Column	: Inertsil ODS-SP (5 μ m, 150 \times 4.6 mm I.D.)		
Column Cat. No.:			2. Propyl p-hydroxybenzoate (0.05 mg/mL)
Eluent	: A) CH ₃ CN B) 2.1 % CH ₃ COOH A/B = 40/60 ,v/v		
Flow rate	: 0.6 mL/min		
Col. Temp.	: 30 °C		
Detection	: PDA 254 nm (GL-7452A PDA detector)		
Injection Vol.	: 20 μ L		
Sample	: Glycyrrhiza Standard		