


# InertSearch™ for LC

Inertsil™ Applications

## Analysis of 16 kinds of Aldehydes (Detected by Pre-Column Method with DNPH)

Data No. LB557-0848



### Conditions

System	: GL7700 HPLC system
Column	: InertSustain PFP (5 $\mu$ m, 250 x 4.6 mm I.D.)
Column Cat. No.	: 5020-87731
Eluent	: A) CH <sub>3</sub> OH/CH <sub>3</sub> CN=9/1, v/v B) H <sub>2</sub> O A/B = 40/60 – 40 min – 53/47 – 20 min – 75/25 – 0.1 min – 40/60, v/v
Flow rate	: 1.0 mL/min
Col. Temp.	: 40 °C
Detection	: UV 360 nm (UV7750 UV Detector)
Injection Vol.	: 10 $\mu$ L
Sample	: Standard

### Analyte:

1. DNPH-Formaldehyde
  2. DNPH-Acetaldehyde
  3. DNPH-Acetone
  4. DNPH-Acrolein
  5. DNPH-Propionaldehyde
  6. DNPH-Crotonaldehyde
  7. DNPH-*n*-Butyraldehyde
  8. DNPH-Isobutyraldehyde
  9. DNPH-Isovaleraldehyde
  10. DNPH-*n*-Valeraldehyde
  11. DNPH-Benzaldehyde
  12. DNPH-Hexanal
  13. DNPH-*o*-Tolualdehyde
  14. DNPH-*m*-Tolualdehyde
  15. DNPH-*p*-Tolualdehyde
  16. DNPH-2,5-dimethylbenzaldehyde
- Conc. 250  $\mu$ g/L each