


## Analysis of Norgestrel

Data No. LA191-0000



Column : Inertsil ODS-3V  
5 $\mu$ m  
250  $\times$  4.6mm I.D.

Solutes : 1. Ethynylestradiol(8 $\mu$ g/mL)  
2. Norgestrel(83 $\mu$ g/mL)  
3. Biphenyl(Diphenyl)(7 $\mu$ g/mL)

Column Cat. No. :

Eluent : CH<sub>3</sub>CN/H<sub>2</sub>O=55/45

Flow rate : 1.2 mL/min.

Col. Temp. : 30°C

Detection : UV 281 nm

Injection Vol. : 20 $\mu$ L

Sample : 1. Ethynylestradiol(8 $\mu$ g/mL)

2. Norgestrel(83 $\mu$ g/mL)

3. Biphenyl(7 $\mu$ g/mL)